

Scan and Login website
for more information



DTE®

FDA CE 0197



TIP BOOK



Contents

Scaling

Model

Page

G1		1
G2		2
G3		3
G4		4
G5		5
G6		6

Dimensions Unit: mm

Model

Page

G7		7
G8		8
G9		9
G10		10
G11		11

Periodontics

Model		Page
P1		12
P2L		13
P2LD		14
P2R		15
P2RD		16
P3		17

Dimensions Unit: mm

Model		Page
P3D		18
P4		19

Endodontics

E1		20
E2		21
E3		22

Endodontics

Model		Page
E 3D		23
E 4		24
E 4D		25
E 5		26

Dimensions Unit: mm

Model		Page
E 5D		27
P 4D		28
E 8		29
E 9		30
E 10		31
E 10D		32

Endodontics




Model		Page
E11		33
E11D		34
E14		35
E15		36

Cavity preparation




Dimensions Unit: mm

Model		Page
SB1		37
SB2		38
SB3		39
SBL		40
SBR		41

Scaling

Model		Page
GK1		42
GK2		43
GK3		44
GK4		45
GK5		46

Dimensions Unit: mm

Model		Page
GK6		47
GK7		48
GC1		49
GC2		50
Periodontics		
PC1		51

Scaling

Model		Page
A1		52
A2		53


Bone surgery

US1		54
US1L		55
US1R		56

Dimensions Unit: mm



Model		Page
US2		57
US3		58
US4		59
US5		60
US6		61

Sinus lifting

Model		Page
UL1		62
UL2		63
UL3		64
UL4		65
UL5		66

Endodontics

Dimensions Unit: mm

Model		Page
UE1		67
UE2		68
UE3		69
UE4		70
UC1		71

Exelcymosis

Implantation

Model		Page
UI1		72
UI2		73
UI7		74
UI8		75
UI9		76




Periodontal surgery

Dimensions Unit: mm



Model		Page
UP1		77
UP2		78
UP3		79
UP4		80
UP5		81
UP6		82
UP7		83

G1 (GD1/GS1)

Removal of supragingival deposits in all quadrants.






MODE: G
POWER





IRRIGATION

G2(GD2/GS2)

Removal of heavy supragingival deposits.
Apply flat end to surface of teeth.






MODE: G
POWER





IRRIGATION

G3(GD3/GS3)

Removal of supragingival deposits in all quadrants, including the interproximal and sulcus areas.






MODE: G
POWER





IRRIGATION

G4(GD4/GS4)

Recommended for the treatment of interproximal spaces and for supragingival scaling.






MODE: G
POWER





IRRIGATION

G5(GD5/GS5)

Recommended for treating simple cases and gross supragingival scaling.






MODE: G
POWER





IRRIGATION

G6(GD6/GS6)

Recommended for removing voluminous supragingival deposits.
Apply flat end to surface of teeth.



MODE: G
POWER





IRRIGATION



G7(GD7/GS7)

Removal of crowns.

Apply the instrument onto surface and activate. Increase pressure until vibrations can no longer be heard and maintain for few seconds.



MODE: G
POWER




IRRIGATION



G8(GD8/GS8)

Removal of posts.

Ultrasonic high efficiency for the removal of the difficult prosthetic parts.





MODE: G
POWER





IRRIGATION

G9(GD9)

Removal of supragingival calculus, interdental calculus and calculus at the neck of the teeth.






MODE: G
POWER





IRRIGATION

G10(GD10)

Removal of supragingival deposits in all quadrants, including the interproximal and sulcus areas.





MODE: G
POWER





IRRIGATION

G11 (GD11/GK11)

Diamond-coated (40µm) instrument for polishing the treatment surface of teeth in interproximal areas, without damaging the adjacent teeth during the orthodontics treatment.






MODE: G
POWER





IRRIGATION

P1 (PD1/PS1)

Slim and sharp, recommended for cleaning and irrigating of subgingival deposits.






MODE: P
POWER





IRRIGATION

P2L(PD2L)

Left-angled, used to remove calculus from very narrow inter-root spaces and furcation.






MODE: P
POWER





IRRIGATION

P2LD(PD2LD)

Left-angled, diamond-coated(40µm) instrument for root planing.





MODE: P
POWER





IRRIGATION

P2R(PD2R)

Right-angled, used to remove calculus from very narrow inter-root spaces and furcation.





MODE: P
POWER





IRRIGATION

P2RD (PD2RD)

Right-angled, diamond-coated(40µm) instrument for root planing.





MODE: P
POWER





IRRIGATION

P3(PD3/PS3)

Recommended for cleaning and irrigating of periodontal deep pockets.






MODE: P
POWER





IRRIGATION

P3D(PD3D/PS3D)

Diamond-coated(40 μ m) instrument for root planning after the periodontal flap surgery. It's also applied to edge planning of artificial crown and furcation expanding.






MODE: P
POWER





P4(PD4/PS4)

Recommended for cleaning and irrigating of periodontal shallow pockets.



MODE: P
POWER




IRRIGATION

E1 (ED1/ES1)

120° holder, used to hold the file diameter $\Phi 0.7\text{mm}$ or $\Phi 0.8\text{mm}$.

Recommended for the cleaning and irrigating of the root canal, usually used for the anterior teeth.





Use only the provided flat wrench to tighten the file holders on the handpiece, as well as for gently tightening the files or instruments in the holder. Do not over tighten.

Do not tighten the holder nut when no file or instrument is installed as this may damage it.

Do not mix-install the holder with $\Phi 0.7\text{mm}$ and $\Phi 0.8\text{mm}$ file, the $\Phi 0.8\text{mm}$ file will not be able to be installed after a $\Phi 0.7\text{mm}$ file installed on the holder.

MODE: E

POWER




IRRIGATION

E2(ED2/ES2)

95° holder, used to hold the file diameter $\Phi 0.7\text{mm}$ or $\Phi 0.8\text{mm}$.

Recommended for cleaning and irrigation of the root canal, usually used for the molar teeth.





Use only the provided flat wrench to tighten the file holders on the handpiece, as well as for gently tightening the files or instruments in the holder. Do not over tighten.

Do not tighten the holder nut when no file or instrument is installed as this may damage it.

Do not mix-install the holder with $\Phi 0.7\text{mm}$ and $\Phi 0.8\text{mm}$ file, the $\Phi 0.8\text{mm}$ file will not be able to be installed after a $\Phi 0.7\text{mm}$ file installed on the holder.

MODE: E

POWER




IRRIGATION

E3(ED3/ES3)



Recommended for melting gutta-percha.

Immerse the instrument in eugenol before inserting it in the canal.

Slightly press the instrument against the cone of gutta-percha and activate the scaler to condense until the canal is filled.



MODE: E
POWER




IRRIGATION



E3D (ED3D/ES3D)

Diamond-coated (40µm) instrument for enlarging the root canal and access to files broken in the root canal.

Don't touch the broken file when enlarging the root canal in case putting it deeper in the canal.





MODE: E
POWER





IRRIGATION

E4(ED4/ES4)

Used to remove the root filling obstructions during the root canal retreatment. The length of tip slender part is (22mm).




MODE: E
POWER





IRRIGATION

E4D (ED4D/ES4D)

Diamond-coated (40 μ m) instrument for removing the calcification and bad filling material in the root canal during the root canal retreatment. The length of the diamond-coated on tip is (15mm).





MODE: E
POWER




IRRIGATION

E5 (ED5/ES5)


Used to remove the root filling obstructions during the root canal retreatment. The length of tip slender part is (16mm).



1





2



3



MODE: E
POWER




IRRIGATION

E5D (ED5D/ES5D)


Diamond-coated (40 μ m) instrument for removing the calcification and bad filling material in the root canal during the root canal retreatment. The length of the diamond-coated on tip is (10mm).



1





2



3

MODE: E
POWER




IRRIGATION



P4D (PD4D/PS4D)

Diamond-coated (40 μ m) instrument for root canal orientation. It can also be applied to remove the calcifications located at the 1/3 part of root canal.

Don't press too much during the treatment to avoid hurting the root canal.



MODE: E
POWER




IRRIGATION



E8(ED8/ES8/EK8)

Recommended for holding different types of dental bur to realize a variety of treatment on teeth.
Usually used for the anterior teeth.

Special burs used for dental implant is applicable.




MODE: E
POWER





IRRIGATION

E9(ED9)

Recommended for holding different types of dental bur to realize a variety of treatment on teeth.
Usually used for the molar teeth.
Special burs used for dental implant is applicable.





MODE: E
POWER




IRRIGATION

E10(ED10/ES10)


Used to polish the root canal in the retrograde preparation of root canals. The length of tip slender part is (4.5mm).



1





2



3





MODE: E
POWER





IRRIGATION

E10D(ED10D/ES10D)

Diamond-coated(40 μ m) instrument for high efficiency root planning in the retrograde preparation of root canals. The length of the diamond-coated on tip is (3.3mm).





MODE: E
POWER





IRRIGATION

E11(ED11/ES11)

Used to polish the root canal in the retrograde preparation of root canals. The length of tip slender part is (3.5mm).




MODE: E
POWER



IRRIGATION

E11D(ED11D/ES11D)

Diamond-coated(40 μ m) instrument for high efficiency root planning in the retrograde preparation of root canals. The length of the diamond-coated on tip is (2.2mm).



1





2



3

MODE: E
POWER




IRRIGATION

E14(ED14/ES15)



Instrument for the removal of broken files inside the root canal. The length of tip slender part is (22mm).

If possible place the patient's head in a position where the root canal is horizontal with a downward inclination. Rotate the tip of the instrument counterclockwise around the broken part until it is picked out from the root canal.

To avoid push the broken file deeper into the root canal, do not use pressure to the instrument in the axial direction.




MODE: E
POWER



IRRIGATION

E14D(ED14D/ES14D)


Diamond-coated(40 μ m) instrument for removing the calcification in the root canal during the root canal treatment. It is used for root canal cleaning and root canal enlargement. The length of the diamond-coated on tip is (15mm).



1





2



3



MODE: E
POWER




IRRIGATION

E15(ED15/ES15)



Instrument for the removal of broken files inside the root canal. The length of tip slender part is (16mm).

If possible place the patient's head in a position where the root canal is horizontal with a downward inclination. Rotate the tip of the instrument counterclockwise around the broken part until it is picked out from the root canal.

To avoid push the broken file deeper into the root canal, do not use pressure to the instrument in the axial direction.





MODE: E
POWER




IRRIGATION

E15D(ED15D/ES15D)


Diamond-coated(40 μ m) instrument for removing the calcification in the root canal during the root canal treatment. It is used for root canal cleaning and root canal enlargement. The length of the diamond-coated on tip is (10mm).



1





2



3

MODE: E
POWER








IRRIGATION



SB1 (SBD1)

Diamond-coated(85µm) instrument for removing the caries on occlusal surface of teeth and dental neck.

Place the instrument on the pit and move slowly into it with light pressure.






MODE: P
POWER





IRRIGATION

SB2 (SBD2)

Diamond-coated(85µm) instrument for removing the caries on Mesial surface of the adjacent teeth.






MODE: P
POWER





IRRIGATION

SB3 (SBD3)

Diamond-coated(85µm) instrument for removing the caries on Distal surface of the adjacent teeth.






MODE: P
POWER





IRRIGATION

SBL(SBDL)

Diamond-coated(85µm) instrument for removing the dental caries, it will not hurt the adjacent teeth from right to left 45 degree.






MODE: P
POWER





IRRIGATION

SBR(SBDR)

Diamond-coated(85 μ m) instrument for removing the dental caries, it will not hurt the adjacent teeth from left to right 45 degree.






MODE: P
POWER





IRRIGATION

GK1

Removal of supragingival deposits in all quadrants.






POWER





IRRIGATION

GK2

Removal of supragingival heavy calculus and plaque.






POWER





IRRIGATION

GK3

Removal of supragingival calculus, interdental calculus and calculus at the neck of the teeth.






POWER





IRRIGATION

GK4

Removal of supragingival deposits, including the interproximal and sulcus areas.






POWER





IRRIGATION

GK5

Removal of subgingival calculus.








POWER



IRRIGATION




GK6

Left-angled, removal of subgingival calculus.





GK7

Right-angled, removal of subgingival calculus.






POWER





IRRIGATION

GC1

Removal of supragingival deposits in all quadrants.






MODE: G
POWER





IRRIGATION

GC2

Removal of supragingival calculus, interdental calculus and calculus at the neck of the teeth.






MODE: G
POWER





IRRIGATION

PC1

Removal of subgingival calculus, including the interproximal and sulcus areas.






MODE: P
POWER





IRRIGATION

A1

Removal of supragingival deposits, including the interproximal and sulcus areas.




POWER





IRRIGATION

A2



Removal of supragingival deposits in all quadrants, including neck and adjacent part of the teeth.



A2







POWER



IRRIGATION


US1

High effectiveness osteotomy of large bone sections during maxillofacial surgery, it can also be used for exodontia exodontics.



MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

US1L


Left angle 90°, horizontal osteotomy technique in maxilla and mandible.



1




2



3

MODE: BONE






POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)



IRRIGATION

US1R

Right angle 90°, horizontal osteotomy technique in maxilla and mandible.




1

2

3

MODE: BONE





POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)



IRRIGATION


US2

High effectiveness osteotomy technique in maxilla and mandible(ridge expansion, corticotomy technique, bone block grafting).



MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

US3


Osteotomy: osteotome of great precision in anatomically thin structures.



1




2



3

MODE: BONE





POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)



IRRIGATION


US4

Universal osteoplasty: periodontal osteotomy, bone chips harvesting, inflammatory tissue removal (cyst, etc.).



MODE: BONE



POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

US5


High efficiency bone osteoplasty: bone remodelling and harvesting of bone chips.



1




2



3

MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

US6

Osteotomy: osteotome of great precision in anatomically thin structures (ridge expansion, corticotomy technique).




MODE: BONE
POWER: Special




UL1

Tip angle 120°, separation of the sinus membrane in internal zones, non-cutting elevator of the sinus membrane.






MODE: ROOT

POWER: Endo(Maximum power permitted)




UL2

Schneiderian membrane separation from bony walls: separation of the sinus membrane.



MODE: ROOT





POWER: Endo(Maximum power permitted)



IRRIGATION

UL3

Diamond-coated(100µm) instrument for micrometric osteotomy or osteoplasty: non-traumatic, to finalize the osteotomy or osteoplasty on thin bone and/or near delicate anatomic structures.




1

2

MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

UL4

Diamond-coated(100µm) instrument for micrometric osteotomy: to finalize the osteotomy in proximity of soft tissue (sinus membrane, vessel, alveolar nerve).




1



2

MODE: BONE


POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)



IRRIGATION


UL5

Tip angle 95°, separation of the sinus membrane in internal zones, non-cutting elevator of the sinus membrane.





MODE: ROOT

POWER: Endo(Maximum power permitted)




UE1


Diamond-coated(40 μ m) instrument for apical root debridement: diamond-coated instrument for efficient root planning. The length of the diamond-coated on tip is (3.3mm).



1




2



3

MODE: ROOT





POWER: Endo(Maximum power permitted)



IRRIGATION


UE2

Used for gentle canal cleaning. The length of tip slender smooth part is (4.5mm).



MODE: ROOT


POWER: Endo(Maximum power permitted)




IRRIGATION

UE3

Diamond-coated(40µm) instrument for apical root debridement: diamond-coated instrument for efficient root planning. The length of the diamond-coated on tip is (2.2mm).



1




2



3

MODE: ROOT



POWER: Endo(Maximum power permitted)




IRRIGATION

UE4


Used for gentle canal cleaning. The length of tip slender smooth part is (3.5mm).



1




2



3

MODE: ROOT



POWER: Endo(Maximum power permitted)



IRRIGATION


UC1

Used to cut off the ankylosis and root fraction techniques.



MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

UI1


Ø1.6mm implantation site preparation insert. The working length of the tip is (9mm).



1




2



3

MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

UI2


Ø2mm implantation site preparation insert. The hole at the center of the top for water spray maximum reduce the heat.



1




2



3

MODE: BONE



POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

UI7


Diamond-coated(85µm) instrument for finalizing the implantation site preparation close to the alveolar nerve.



1




2



3

MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

U18


Ø1.6mm implantation site preparation insert. The working length of the tip is (15mm).



1




2



3

MODE: BONE




POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)




IRRIGATION

U19


Ø2.8mm implantation site preparation insert . The hole at the center of the top for water spray maximum reduce the heat.



1




2



3

MODE: BONE



POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)



IRRIGATION

UP1


Recommended for periodontal gentle curetting scaling.



1




2



3




MODE: ROOT
POWER: PERIO




IRRIGATION

UP2


Inflammatory tissue removal and fractured root apex extraction.



1




2



3





MODE: ROOT
POWER: PERIO




IRRIGATION

UP3

Recommended for angled periodontal gentle curetting scaling.







MODE: ROOT
POWER: PERIO




IRRIGATION

UP4

Diamond-coated(40µm) instrument for root debridement and root planning.






MODE: ROOT
POWER: PERIO




IRRIGATION

UP5

Recommended for root surface micro-smoothing.




1



2




MODE: ROOT
POWER: PERIO




IRRIGATION

UP6

Diamond-coated(40µm) instrument for micro-osteoplasty: interproximal osteoplasty and root planning.




1



2

MODE: ROOT


POWER: Quality1, Quality2, Quality3
(Cortical/Spongious)



IRRIGATION


UP7

Diamond-coated(40 μ m) instrument for root canal orientation and removal of the calcifications located at the 1/3 part of root canal, it can also be used for micro-osteoplasty.



MODE: ROOT

POWER: Endo(Maximum power permitted)



RECOMMENDED POWER ON WOODPECKER SCALER

Tip Model	WOODPECKER Scaler Model	UDS-E (LED) UDS-L (LED) UDS-A (LED) UDS-P (LED)	UDS-J2 (LED)	UDS-J UDS-B UDS-K (LED)	UDS-N3 (LED)	UDS-N1 UDS-N2 (LED) UDS-N4	Compatible Scaler Brand
Scaling							
G1	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH	Compatible with WOODPECKER & EMS Ultrasonic Scalers	
G2	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G3	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G4	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G5	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G6	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G7	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G8	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G9	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G10	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
G11	1-10 (G)	1-10	1-9	LOW-HIGH	LOW-HIGH		
Periodontics							
P1	1-10 (P)	1-7	1-6	LOW-MID	LOW-MID	Compatible with WOODPECKER & EMS Ultrasonic Scalers	
P2L	1-3 (P)	1-2	1-2	LOW	LOW		
P2LD	1-2 (P)	1	1	LOW	LOW		
P2R	1-3 (P)	1-2	1-2	LOW	LOW		
P2RD	1-2 (P)	1	1	LOW	LOW		
P3	1-6 (P)	1-4	1-3	LOW-MID	LOW-MID		
P3D	1-6 (P)	1-4	1-3	LOW-MID	LOW-MID		
P4	1-6 (P)	1-4	1-3	LOW-MID	LOW-MID		

RECOMMENDED POWER ON WOODPECKER SCALER

Tip Model	WOODPECKER Scaler Model	UDS-E (LED) UDS-L (LED) UDS-A (LED) UDS-P (LED)	UDS-J2 (LED)	UDS-J UDS-B UDS-K (LED)	UDS-N3 (LED)	UDS-N1 UDS-N2 (LED) UDS-N4	Compatible Scaler Brand
Endodontics							
E1	1-3 (E)	—	—	LOW	—	Compatible with WOODPECKER & EMS Ultrasonic Scalers	
E2	1-3 (E)	—	—	LOW	—		
E3	1-6 (E)	—	—	LOW	—		
E3D	1-3 (E)	—	—	LOW	—		
E4	1-6 (E)	—	—	LOW	—		
E4D	1-3 (E)	—	—	LOW	—		
E5	1-6 (E)	—	—	LOW	—		
E5D	1-3 (E)	—	—	LOW	—		
P4D	1-6 (E)	—	—	LOW	—		
E8	1-10 (E)	—	—	LOW	—		
E9	1-10 (E)	—	—	LOW	—		
E10	1-6 (E)	—	—	LOW	—		
E10D	1-6 (E)	—	—	LOW	—		
E11	1-6 (E)	—	—	LOW	—		
E11D	1-6 (E)	—	—	LOW	—		
E14	1-3 (E)	—	—	LOW	—		
E15	1-3 (E)	—	—	LOW	—		

RECOMMENDED POWER ON WOODPECKER SCALER

Tip Model	WOODPECKER Scaler Model	UDS-E (LED) UDS-L (LED) UDS-A (LED) UDS-P (LED)	UDS-J2 (LED)	UDS-J UDS-B UDS-K (LED)	UDS-N3 (LED)	UDS-N1 UDS-N2 (LED) UDS-N4	Compatible Scaler Brand
SB1	1-10 (P)	1-7	1-6	LOW-MID	LOW-MID	Compatible with WOODPECKER & EMS Ultrasonic Scalers	
SB2	1-10 (P)	1-7	1-6	LOW-MID	LOW-MID		
SB3	1-10 (P)	1-7	1-6	LOW-MID	LOW-MID		
SBL	1-10 (P)	1-7	1-6	LOW-MID	LOW-MID		
SBR	1-10 (P)	1-7	1-6	LOW-MID	LOW-MID		

[NOTE] : "G" for the working mode of "Scaling"; "P" for the working mode of "Periodontics"; "E" for the working mode of "Endodontics"; "—" for "not suitable for such model of scaler".

RECOMMENDED POWER ON DTE SCALER

Tip Model	DTE Scaler Model Power	D7 (LED) D5 (LED)	D2 LED	D3 (LED)	D1	V3 (LED)	V1 V2 (LED)	Compatible Scaler Brand
		Scaling						
GD1		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	Compatible with DTE & Satelec Ultrasonic Scalers
GD2		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD3		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD4		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD5		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD6		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD7		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD8		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD9		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD10		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
GD11		1-10(G)	1-11	1-10(G)	1-9	LOW-HIGH	LOW-HIGH	
Periodontics								
PD1		1-10(P)	1-8	1-10(P)	1-6	LOW-MID	LOW-MID	Compatible with DTE & Satelec Ultrasonic Scalers
PD2L		1-3(P)	1-3	1-3(P)	1-2	LOW	LOW	
PD2LD		1-2(P)	1-2	1-2(P)	1	LOW	LOW	
PD2R		1-3(P)	1-3	1-3(P)	1-2	LOW	LOW	
PD2RD		1-2(P)	1-2	1-2(P)	1	LOW	LOW	
PD3		1-6(P)	1-5	1-6(P)	1-3	LOW-MID	LOW-MID	
PD3D		1-6(P)	1-5	1-6(P)	1-3	LOW-MID	LOW-MID	
PD4		1-6(P)	1-5	1-6(P)	1-3	LOW-MID	LOW-MID	

RECOMMENDED POWER ON DTE SCALER

Tip Model	DTE Scaler Model	D7 (LED) D5 (LED)	D2 LED	D3 (LED)	D1	V3 (LED)	V1 V2 (LED)	Compatible Scaler Brand
	Power							
Endodontics								
ED1		1-3 (E)	—	—	—	LOW	—	Compatible with DTE & Satelec Ultrasonic Scalers
ED2		1-3 (E)	—	—	—	LOW	—	
ED3		1-6 (E)	—	—	—	LOW	—	
ED3D		1-3 (E)	—	—	—	LOW	—	
ED4		1-6 (E)	—	—	—	LOW	—	
ED4D		1-3 (E)	—	—	—	LOW	—	
ED5		1-6 (E)	—	—	—	LOW	—	
ED5D		1-3 (E)	—	—	—	LOW	—	
PD4D		1-6 (E)	—	—	—	LOW	—	
ED8		1-10 (E)	—	—	—	LOW	—	
ED9		1-10 (E)	—	—	—	LOW	—	
ED10		1-6 (E)	—	—	—	LOW	—	
ED10D		1-6 (E)	—	—	—	LOW	—	
ED11		1-6 (E)	—	—	—	LOW	—	
ED11D		1-6 (E)	—	—	—	LOW	—	
ED14		1-3 (E)	—	—	—	LOW	—	
ED15		1-3 (E)	—	—	—	LOW	—	

RECOMMENDED POWER ON DTE SCALER

Tip Model	DTE Scaler Model	D7 (LED) D5 (LED)	D2 LED	D3 (LED)	D1	V3 (LED)	V1 V2 (LED)	Compatible Scaler Brand
	Power							
Cavity Preparation								
SBD1		1-10(P)	1-8	1-10(P)	1-6	LOW-MID	LOW-MID	Compatible with DTE & Satelec Ultrasonic Scalers
SBD2		1-10(P)	1-8	1-10(P)	1-6	LOW-MID	LOW-MID	
SBD3		1-10(P)	1-8	1-10(P)	1-6	LOW-MID	LOW-MID	
SBDL		1-10(P)	1-8	1-10(P)	1-6	LOW-MID	LOW-MID	
SBDR		1-10(P)	1-8	1-10(P)	1-6	LOW-MID	LOW-MID	

[NOTE] : "G" for the working mode of "Scaling"; "P" for the working mode of "Periodontics"; "E" for the working mode of "Endodontics"; "—" for "not suitable for such model of scaler".

MODE OF ULTRASURGERY AND OTHER SPECIAL TIPS

	Tip Model	Compatible Brand
Scaling	GS1/GS2/GS3/GS4/GS5/GS6/GS7/GS8	Compatible with Sirona Scalers
Periodontics	PS1/PS3/PS3D/PS4	
Endodontics	ES1/ES2/ES3/ES3D/ES4/ES4D/ES5/ES5D/ PS4D/ES8/ES10/ES10D/ES11/ES11D/ES14 /ES15	
Scaling	GK1/GK2/GK3/GK4/GK5/GK6/GK7/GK11	Compatible with Kavo Scalers
Endodontics	EK8	
Scaling	GC1/GC2	
Periodontics	PC1	
Scaling	A1/A2	Compatible with Amdent Scalers
Bone surgery	US1/US1L/US1R/US2/US3/US4/US5/US6	Compatible with WOODPECKER & Mectron Ultrasurgery
Sinus lifting	UL1/UL2/UL3/UL4/UL5	
Endodontics	UE1/UE2/UE3/UE4	
Exelcymosis	UC1	
Implantation	UI1/UI2/UI7/UI8/UI9	
Periodontal surgery	UP1/UP2/UP3/UP4/UP5/UP6/UP7	



www.glwoodpecker.com

 Guilin Woodpecker Medical Instrument Co., Ltd.
Information Industrial Park, National High-Tech
Zone, Guilin, Guangxi, 541004 P. R. China

Tel:
Europe Sales Dept.: +86-773-5873196, +86-773-2125222
North America, South America &
Oceania Sales Dept.: +86-773-5873198, +86-773-2125123
Asia & Africa Sales Dept.: +86-773-5855350, +86-773-2125896

Fax: +86-773-5822450
E-mail: woodpecker@mailgl.cn, sales@glwoodpecker.cn
Website: <http://www.glwoodpecker.com>

 Wellkang Ltd (www.CE-Marking.eu)
29 Harley St., LONDON, W1G 9QR, UK

ZMN/WI-09-036 1. 1

